



LUMEN DYNAMICS

PUTTING YOU IN CONTROL

An Excelitas Technologies® Company



CE Declaration of Conformity

Manufacturer's Name: Lumen Dynamics Group Inc.
Manufacturer's Address: 2260 Argentia Road, Mississauga, Ontario L5N 6H7 Canada

Authorized EU Representative: Qioptic Photonics GmbH & Co. KG
Hans-Riedl-Str. 9, D-85622 Feldkirchen, Germany
Tel: (+49) 89-255 458-0/ (+49) 89-255 458-111

Declares that the product: **Equipment Type:** Microscopy Illumination Device
Trade Name/Model No.: X-Cite / XLED1

Conforms to the following Council Directives:

Low Voltage	2014/35/EU
EMC	2014/30/EU
WEEE	2012/19/EU
RoHS	2011/65/EU

Conforms to the following Standards:

Product Safety

CAN/CSA C22.2 No 61010-1-04 (2nd Edition)
UL 61010-1:2004 (2nd Edition)

Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements

IEC 61010-1:2010 (Third Edition)

Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements

IEC 62471:2006 (1st Edition)

Photobiological Safety of Lamps and Lamp Systems

IEC/TR 62471-2:2009

Technical Report: Guidance on manufacturing Requirements relating to Non-Laser Optical Radiation Safety

Electromagnetic Compatibility

EN 61326-1:2013

Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements, Part 1: General Requirements

CISPR 11:2009 + A1:2010 / EN55011:2009 + A1:2010, Class A, Group 1

Industrial, Scientific and Medical Equipment

FCC Part 15, Subpart B, Class A

Unintentional Radiators

I, the undersigned, hereby declare that the equipment specified above conforms to the above directives and standards specified.

Signature 

Date MAR 2 1 2017

Full Name: Jim Farrell
Position: QA, QMS & EMS Manager



www.excelitas.com/x-cite
x-cite@excelitas.com

www.excelitas.com/omnicure
omnicure@excelitas.com

Lumen Dynamics Group Inc. 2260 Argentia Road, Mississauga, Ontario L5N 6H7 CANADA
Telephone: +1 905 821-2600 • Toll Free (USA and CAN): +1 800 666-8752 • Fax: +1 905 821-2055